


# Investigation of Converging and Collimated Beam Instrument Geometry on Specular Gloss Measurements



Joanne Zwinkels\*, Eric Cote\* and John Morgan

\* National Research Council Canada

11<sup>th</sup> Biennial USNC&CNC/CIE technical day, NIST Gaithersburg, 3 Oct. 2017




# Outline



- Gloss Measurement
  - Principles
  - Standardization
- Current situation
  - Repeatability & Reproducibility
  - Q. What is impact of beam geometry: converging or collimated?
- NRC gloss study
  - NRC Reference Goniospectrophotometer (GSP)
  - Comparison samples
  - Results for different beam geometries
- Conclusions

# Introduction


**Specular gloss measurements are used by industry to describe the surface shininess or roughness of materials by detecting the mirror-like (specular) reflection.**



# Standardization of Gloss Measurements: **Source**




# Standardization of Gloss Measurements: **Detector**



# Standardization of Gloss Measurements: **Standard**

Depends on Application

NOTE: also depends on Specification




Application	Specification	Primary Standard
PAPER	ISO 8254-2	$G_s = \text{black glass}$ $(n = 1.567) = 100$

Application	Specification	Primary Standard
PAINT	ASTM D523 ISO 2813	$G_s = \text{black glass}$ $(n = 1.567) = 100$
PLASTICS	ASTM D2457	$G_s = \text{perfect mirror}$ $= 1000$
PAPER	TAPPI T653 ISO 8254-3 ISO 8254-1	$G_s = \text{black glass}$ $(n_D = 1.540) = 100$

# Measurement of Specular Gloss: Definition


Measured gloss depends on specular angle

NOTE: also depends on other geometric conditions



$$\text{Gloss} = \frac{I_r}{I_0} = \frac{\text{received light (through detector field stop)}}{\text{incident light (through source field stop)}}$$

# Standardization of Gloss Measurements: **Geometry**



Converging  
beam

or  
Collimated  
(Parallel)  
beam

**ASTM D523 specifications for source and receptor apertures cf. NRC Goniospectrophotometer**

Apertures	In Plane (°)		Normal to Plane (°)	
	ASTM	NRC	ASTM.	NRC
Source	0.75 ± 0.25	0.75 ± 0.01	3.0 ± 0.5	3.00 ± 0.01
20° receptor	1.80 ± 0.05	1.81 ± 0.01	3.6 ± 0.1	3.62 ± 0.01
60° receptor	4.40 ± 0.25	4.42 ± 0.01	11.7 ± 0.2	11.75 ± 0.01
85° receptor	4.00 ± 0.25	4.02 ± 0.01	6.0 ± 0.3	6.02 ± 0.01
Specular angle	20, 60, 85 ±1	20, 60, 85 ±0.05		

Depends on  
Specification

Depends on  
Beam  
Geometry




# Standardization of Gloss Measurements: Summary

## Spectral Conditions:

spectral nature of the illumination  
spectral responsivity of the detection system

## Geometric Conditions:

Angle of incidence  
Source and receptor aperture dimensions



## Gloss Scale:

Definition of primary gloss standard

## Other Geometric Conditions:

Convergence angle:

**converging or collimated beam**

**Impact not been studied**

# Gloss – Standard Test Methods: 75° and 20° Geometries

## 75° Gloss: Paper and Board

- ISO 8254-1 *Measurement of specular gloss – Part 1: 75° gloss with a converging beam: TAPPI method*
- ISO 8254-2 *Measurement of specular gloss – Part 2: 75° gloss with a parallel beam: DIN method*
- TAPPI T480 *Specular gloss of paper and paperboard at 75°*

## 20° Gloss: Paper and Board

- ISO 8254-3 *Measurement of specular gloss – Part 3: 20° gloss with a converging beam: TAPPI method*
- TAPPI T653 *Specular gloss of paper and paperboard at 20°*

## 20° Gloss: Paint and General Material

- ISO 2813/ ASTM D523 *Standard test method for specular gloss*

Instrument geometry: converging beam; collimated beam

Primary gloss standard refractive index: 1.540 (TAPPI); 1.567 (DIN)

# Glossmeters: Repeatability & Reproducibility

## Manufacturer specifications (20°/60°/85° geometry)

- Repeatability:  $\pm 0.1$  to 0.2 GU
- Reproducibility:  $\pm 0.5$  GU

## c.f. Standards specifications (ASTM D523/ ISO 2813)

- Repeatability:  $\pm 0.8$  to 1.7 GU
- Reproducibility:  $\pm 2.0$  to 7.2 GU

## c.f. In practice (6 commercial glossmeters; 25 samples)\*:

- Repeatability:  $\pm 0.4$  to  $\pm 17.1$  GU
- Reproducibility:  $\pm 0.8$  to 14.3 GU

\* F.B. Leloup et.al, *Repeatability and reproducibility of specular gloss meters in theory and practice*, J. Coat.Technol. Res., Vol. 13 (2016)

# TAPPI 20° Gloss Round Robin (RR) Study (2016)

## Background:


Optical Properties Committee of TAPPI's P&PQ Division requested CTS to conduct a RR study to answer:

- **Q.** Can the technical parameters of TAPPI Method T653: *Specular gloss of paper and paperboard at 20°* be loosened?

## RR Protocol:

- Test Sample: Medium gloss paper (RG01) of 30 specimens (sheets)
- Measure each sheet in upstream and downstream direction
  - e.g. 1<sup>st</sup> meas = 20.7 GU
  - rotate sheet 180°
  - 2<sup>nd</sup> meas = 21.4 GU
  - Reported value = 21.05 GU
- No. of participants: 14 Labs

# CTS Results for TAPPI 20° Gloss RR Study



Grand Mean: 44.9 GU

Standard deviation (SD): 8.7 GU


# NRC Reference Goniospectrophotometer (GSP)

- **Developed for multi-functional purposes:**
  - Gloss at 20,60,85° (ASTM D523 – paints)
  - Gloss at 75° (TAPPI T480 – paper)
  - Gonireflectance measurements (metallic and pearlescent colors)
- **Versatile instrument design:**
  - rapid conversion between converging & collimated beam geometries
- **Spectral and geometric properties of instrument:**
  - measured and compared with standard requirements<sup>1</sup>
- **Uncertainties in gloss measurements have been estimated:<sup>2</sup>**
  - Modelling deviation from standard requirements
  - Measurement intercomparisons (NRC Glossmeter and CTS)
  - Comparison with calculated values


<sup>1</sup> J. Liu, M. Noel and J. Zwinkels, *Appl. Opt.*, **44**, 4631-4638 (2005)

<sup>2</sup> M. Noel, J. Zwinkels and J. Liu, *Appl. Opt.*, **45**, 3712-3720 (2006)

# NRC Reference Goniospectrophotometer (GSP)



# NRC Reference Goniospectrophotometer (GSP)







# Re-configure GSP from collimated to converging beam

## DIN method


### ASTM D523 Configuration



detector stop

## TAPPI method

### ASTM D1223 Configuration



Optional lens assembly (aperture stop)



# NRC Gloss Study: Impact of Beam Geometry


## Paper Samples Tested (from CTS):

### 75° Geometry

<u>ID</u>	<u>Type (# sheets)</u>
GU53	Medium gloss (10)
GT53	High gloss (10)


### 20° Geometry

<u>ID</u>	<u>Type (# sheets)</u>
RG01	Medium gloss (30)



Paper Sample Mount

Vacuum suction plate



Pitch and yaw mounting stage

# GSP Measurement Procedure and Validation (Ex.)

20° Gloss - TAPPI Method; Sample: RG01; Run #1

#1	Meas 1	Meas 2	temp	Hum	#		Meas	STD	Error	temp	Hum	
1	36.065	35.795	27	21	681		NRC High	93.09	0.055	0.04	27	21
2	38.072	34.722	27	21	683		Vitrolite	86.86	0.075	0.07	27	21
3	36.859	35.961	27	21	685							
4	35.317	36.441	27	21	687							
5	36.444	37.588	27	21	689							
6	36.595	36.521	27	21	691							
7	36.302	36.075	27	21	693							
8	36.8	33.925	27	21	696							
9	34.842	34.532	27	21	698							
10	35.678	35.005	27	21	700							
11	37.13	36.431	27	21	702							
12	36.536	36.244	27	21	704							
13	36.071	35.504	27	20	706							
14	36.019	34.963	27	20	708							
15	36.638	35.965	27	21	710		Vitrolite	86.92	0.007	0.13	27	21
16	36.88	35.19	27	21	714							
17	36.39	35.688	27	21	716							
18	35.869	36.379	27	21	718							
19	35.77	34.794	27	21	720							
20	34.723	34.38	27	20	722							
21	36.924	36.5	27	20	724							
22	36.249	36.563	27	20	726							
23	34.657	34.352	27	20	728							
24	35.989	35.797	27	20	730							
25	37.119	35.839	27	20	732							
26	36.32	36.287	27	20	734							
27	35.989	37.23	27	20	736							
28	36.226	35.789	27	20	738							
29	36.671	36.814	27	20	740							
30	36.757	35.697	27	20	742		Vitrolite	87.119	0.0136	0.329	27	20
							NRC High	93.36	0.0066	0.31	27	19

NRC Quality System  
Validation Standards

# GSP Measurements and Validation Procedure

## 20 deg. Gloss Validation

	NRC		DIN		TAPPI	
	NRC-Vitrolite	NRC-High-gloss	RG01 #1	RG01 #2	RG01 #1	RG01 #2
Set #1 of Meas.	86.88	93.26	35.95	34.95	58.85	58.85
			36.40	34.65	59.15	59.15
			36.45	34.10	59.50	57.45
			35.85	32.80	59.25	59.05
Mean	<b>86.88</b>	<b>93.26</b>	37.00	34.85	59.90	58.80
Std.dev.	<b>0.04</b>	<b>0.05</b>	36.55	34.00	58.95	58.40
			36.20	35.20	59.50	58.50
Set #2 of Meas.	86.97	93.28	35.35	36.40	58.70	59.40
			34.65	35.70	58.90	57.85
			35.35	34.15	59.15	59.90
			36.75	34.55	59.60	59.50
Mean	<b>86.97</b>	<b>93.28</b>	36.35	34.30	59.90	59.30
Std.dev.	<b>0.05</b>	<b>0.06</b>	35.80	35.15	58.70	58.75
			35.50	35.20	57.95	59.10
Set #3 of Meas.	86.90	93.24	36.30	35.20	58.55	60.55
			36.05	31.95	58.45	58.85
			36.05	36.70	58.55	59.20
			36.15	34.90	58.55	57.65
Mean	<b>86.90</b>	<b>93.24</b>	35.30	33.85	58.80	59.40
Std.dev.	<b>0.08</b>	<b>0.10</b>	34.55	32.00	56.95	59.50
			36.70	32.30	59.60	58.65
	86.88	93.26	36.40	35.85	59.05	58.05
	86.97	93.28	34.55	33.10	58.40	58.35
	86.90	93.24	35.90	35.50	58.55	57.90
Mean of means	<b>86.92</b>	<b>93.26</b>	36.45	35.80	59.25	60.65
Std.dev.	<b>0.05</b>	<b>0.02</b>	36.30	34.20	59.40	59.65
			36.60	35.70	59.30	59.00
			36.00	34.55	59.10	59.25
			36.75	34.35	59.65	58.70
			36.25	33.40	59.15	59.45
Historical Value,	<b>86.79</b>	<b>93.05</b>	<b>36.02</b>	<b>34.51</b>	<b>58.98</b>	<b>58.96</b>
Stdm	<b>0.02</b>	<b>0.19</b>	<b>0.65</b>	<b>1.22</b>	<b>0.61</b>	<b>0.76</b>
n	<b>113</b>	<b>60</b>				
Degree of Equivalence	<b>0.15%</b>	<b>0.22%</b>				

DoE = 0.2 %

# NRC 75° Gloss Results

Collimated Beam (DIN method)

ID	Run 1 (GU)	Run 2 (GU)
<b>GU53</b>	33.03	31.03
S.D.	±0.75	±0.73
<b>GT53</b>	66.43	65.30
S.D.	±2.00	±2.36
<b>Quartz Standard</b>	92.65	92.65

Converging (TAPPI method)


ID	Run 1 (GU)	Run 2 (GU)
<b>GU53</b>	37.94	38.73
S.D.	±0.82	±0.83
<b>GT53</b>	73.38	73.98
S.D.	±1.94	±1.91
<b>Quartz Standard</b>	94.22	94.22

Statistically significant differences:

**GU53:**  $35.18 \pm 3.74$  GU (c.f. GSP reproducibility:  $\pm 0.78$  GU)

**GT53:**  $69.77 \pm 4.54$  GU (c.f. GSP reproducibility:  $\pm 2.05$  GU)

# NRC Results for TAPPI 20° Gloss RR Study



**X = TAPPI method**

Run 1: 36.02 ± 0.65 GU  
 Run 2: 34.50 ± 1.22 GU



**Y = DIN method**

Run 1: 58.98 ± 0.61 GU  
 Run 2: 58.96 ± 0.76 GU

$\Delta$ Gloss (DIN-TAPPI)  
 = **23.7 GU**

GSP reproducibility:  
 ± 0.6 to 1.2 GU

**c.f. CTS RR Results for TAPPI Method:**  
 Grand Mean: 44.9 GU ± 8.7 GU (14 labs)  
 NRC Mean: 47.1 GU ± 13.7 GU (both methods)

## Conclusions of this NRC gloss study

- For a given specular angle and **given beam geometry**, the reproducibility is  **$\pm 0.6 - \pm 2.0$  GU**
- For both  $20^\circ$  and  $75^\circ$  geometries, for a **change in beam geometry**, the reproducibility is  **$\pm 3.7 - \pm 13.7$  GU**
- For  **$20^\circ$**  geometry, the gloss values **increase** in going from **converging** to **collimated** beam geometry
- For  **$75^\circ$**  geometry, the gloss values **decrease** in going from **converging** to **collimated** beam geometry
- Results indicate that gloss differences observed in recent comparisons may be due to glossmeters with incorrect beam geometry

**For optimum measurement reproducibility:** instrument needs to strictly conform to geometric, spectral and photometric conditions of the specified standard test method, within specified tolerances

**- Including requirements for beam geometry!**

Questions???

**Thank you**

**Joanne Zwinkels**

Principal Research Officer

[Joanne.zwinkels@nrc-cnrc.gc.ca](mailto:Joanne.zwinkels@nrc-cnrc.gc.ca)

[www.nrc-cnrc.gc.ca](http://www.nrc-cnrc.gc.ca)

